
Evaluation of Lots of Dehydrated Mueller-Hinton Broth for Antimicrobial Susceptibility Testing; Proposed Guideline

PLEASE



This proposed document is published for wide and thorough review as the first step in the NCCLS consensus-review process. Please send your comments on scope, approach, and technical and editorial content to the Executive Offices.

Comment period ends
12 March 2002

The subcommittee responsible for this document will assess all comments received by the end of the comment period. Based on this assessment, a new version of the document will be issued. Readers are encouraged to send their comments to the NCCLS Executive Offices, 940 West Valley Road, Suite 1400, Wayne, PA 19087-1898; Fax: (610) 688.0700, or to the following e-mail address: standard@nccls.org



COMMENT

This document provides protocols for evaluating the performance characteristics of production lots of dehydrated Mueller-Hinton broth (dMHB).

A guideline for global application developed through the NCCLS consensus process.



NCCLS...

Serving the World's Medical Science Community Through Voluntary Consensus

NCCLS is an international, interdisciplinary, nonprofit, standards-developing, and educational organization that promotes the development and use of voluntary consensus standards and guidelines within the healthcare community. It is recognized worldwide for the application of its unique consensus process in the development of standards and guidelines for patient testing and related healthcare issues. NCCLS is based on the principle that consensus is an effective and cost-effective way to improve patient testing and healthcare services.

In addition to developing and promoting the use of voluntary consensus standards and guidelines, NCCLS provides an open and unbiased forum to address critical issues affecting the quality of patient testing and health care.

PUBLICATIONS

An NCCLS document is published as a standard, guideline, or committee report.

Standard A document developed through the consensus process that clearly identifies specific, essential requirements for materials, methods, or practices for use in an unmodified form. A standard may, in addition, contain discretionary elements, which are clearly identified.

Guideline A document developed through the consensus process describing criteria for a general operating practice, procedure, or material for voluntary use. A guideline may be used as written or modified by the user to fit specific needs.

Report A document that has not been subjected to consensus review and is released by the Board of Directors.

CONSENSUS PROCESS

The NCCLS voluntary consensus process is a protocol establishing formal criteria for:

- the authorization of a project
- the development and open review of documents
- the revision of documents in response to comments by users
- the acceptance of a document as a consensus standard or guideline.

Most NCCLS documents are subject to two levels of consensus—"proposed" and "approved." Depending on

the need for field evaluation or data collection, documents may also be made available for review at an intermediate (i.e., "tentative") consensus level.

Proposed An NCCLS consensus document undergoes the first stage of review by the healthcare community as a proposed standard or guideline. The document should receive a wide and thorough technical review, including an overall review of its scope, approach, and utility, and a line-by-line review of its technical and editorial content.

Tentative A tentative standard or guideline is made available for review and comment only when a recommended method has a well-defined need for a field evaluation or when a recommended protocol requires that specific data be collected. It should be reviewed to ensure its utility.

Approved An approved standard or guideline has achieved consensus within the healthcare community. It should be reviewed to assess the utility of the final document, to ensure attainment of consensus (i.e., that comments on earlier versions have been satisfactorily addressed), and to identify the need for additional consensus documents.

NCCLS standards and guidelines represent a consensus opinion on good practices and reflect the substantial agreement by materially affected, competent, and interested parties obtained by following NCCLS's established consensus procedures. Provisions in NCCLS standards and guidelines may be more or less stringent than applicable regulations. Consequently, conformance to this voluntary consensus document does not relieve the user of responsibility for compliance with applicable regulations.

COMMENTS

The comments of users are essential to the consensus process. Anyone may submit a comment, and all comments are addressed, according to the consensus process, by the NCCLS committee that wrote the document. All comments, including those that result in a change to the document when published at the next consensus level and those that do not result in a change, are responded to by the committee in an appendix to the document. Readers are strongly encouraged to comment in any form and at any time on any NCCLS document. Address comments to the NCCLS Executive Offices, 940 West Valley Road, Suite 1400, Wayne, PA 19087, USA.

VOLUNTEER PARTICIPATION

Healthcare professionals in all specialties are urged to volunteer for participation in NCCLS projects. Please contact the NCCLS Executive Offices for additional information on committee participation.

Evaluation of Lots of Dehydrated Mueller-Hinton Broth for Antimicrobial Susceptibility Testing; Proposed Guideline

Abstract

This document describes methods for evaluation of production lots of Mueller-Hinton broth by manufacturers of the dehydrated product. Performance characteristics of production lots and broth prepared with the dehydrated product are determined by testing defined organism/antimicrobial agent combinations. The results of testing must conform to defined quality control limit ranges for each combination of antimicrobial agent and ATCC quality control strain. Each production lot should be tested at least against these combinations of antimicrobial agents and quality control strains. Lots falling outside these limits should be retested, and ion contents should be checked to ensure that they are within suggested ranges. Concentration ranges or limits of calcium (20 to 25 mg/L), magnesium (10 to 12.5 mg/L) and zinc (< 3mg/L) are provided. It is expected that media that contain these ranges or limits of cation concentrations will provide results within the defined quality control ranges.

This document will be updated when primary and secondary “standard” lots of Mueller-Hinton broth powder have been selected, so that manufacturers can check production lots against these standard lots of Mueller-Hinton broth media. The studies on which this proposed document is based were designed to test selected class-representative antimicrobial agents against specific quality control strains. As new antimicrobial agents are discovered for which Mueller-Hinton broth dilution quality control ranges are defined by NCCLS and FDA, the primary and secondary “standard” Mueller-Hinton broth media lots should be checked against those agents also. That information will permit the NCCLS Subcommittee on Culture Media to keep future revisions to this document up to date.

NCCLS. *Evaluation of Lots of Dehydrated Mueller-Hinton Broth for Antimicrobial Susceptibility Testing; Proposed Guideline*. NCCLS document M32-P (ISBN 1-56238-447-3). NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA 2001.

THE NCCLS consensus process, which is the mechanism for moving a document through two or more levels of review by the healthcare community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of NCCLS documents. Current editions are listed in the *NCCLS Catalog*, which is distributed to member organizations, and to nonmembers on request. If your organization is not a member and would like to become one, and to request a copy of the *NCCLS Catalog*, contact the NCCLS Executive Offices. Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: exoffice@nccls.org; Website: www.nccls.org

M32-P
ISBN 1-56238-447-3
ISSN 0273-3099

Evaluation of Lots of Dehydrated Mueller-Hinton Broth for Antimicrobial Susceptibility Testing; Proposed Guideline

Volume 21 Number 21

Robert P. Rennie, Ph.D., Chairholder
Arthur L. Barry, Ph.D.
Donald R. Callihan, Ph.D.
Karen Krisher, Ph.D.
Brenda McCurdy, Ph.D.
Roxanne G. Shively, M.S.
Jana M. Swenson, M.M.Sc.



This publication is protected by copyright. No part of it may be reproduced, stored in a retrieval system, transmitted, or made available in any form or by any means (electronic, mechanical, photocopying, recording, or otherwise) without prior written permission from NCCLS, except as stated below.

NCCLS hereby grants permission to reproduce limited portions of this publication for use in laboratory procedure manuals at a single site, for interlibrary loan, or for use in educational programs provided that multiple copies of such reproduction shall include the following notice, be distributed without charge, and, in no event, contain more than 20% of the document's text.

Reproduced with permission, from NCCLS publication M32-P— *Evaluation of Lots of Dehydrated Mueller-Hinton Broth for Antimicrobial Susceptibility Testing; Proposed Guideline* (ISBN 1-56238-447-3). Copies of the current edition may be obtained from NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA.

Permission to reproduce or otherwise use the text of this document to an extent that exceeds the exemptions granted here or under the Copyright Law must be obtained from NCCLS by written request. To request such permission, address inquiries to the Executive Director, NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA.

Copyright ©2001. The National Committee for Clinical Laboratory Standards.

Suggested Citation

(NCCLS. *Evaluation of Lots of Dehydrated Mueller-Hinton Broth for Antimicrobial Susceptibility Testing; Proposed Guideline*. NCCLS document M32-P [ISBN 1-56238-447-3]. NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2001.)

Proposed Guideline

December 2001

ISBN 1-56238-447-3
ISSN 0273-3099

Committee Membership

Area Committee on Microbiology

James H. Jorgensen, Ph.D.
Chairholder

University of Texas Health Center
San Antonio, Texas

Melvin P. Weinstein, M.D.
Vice-Chairholder

Robert Wood Johnson Medical School
New Brunswick, New Jersey

Subcommittee on Culture Media

Robert P. Rennie, Ph.D.
Chairholder

University of Alberta Hospital
Edmonton, Alberta, Canada

Arthur L. Barry, Ph.D.

Clinical Microbiology Institute
Wilsonville, Oregon

Donald R. Callihan, Ph.D.

BD Diagnostic Systems
Sparks, Maryland

Karen Krisher, Ph.D.

Children's Medical Center of Dallas
Dallas, Texas

Brenda McCurdy, Ph.D.

Detroit VA Medical Center
Detroit, Michigan

Roxanne G. Shively, M.S.

Food and Drug Administration
Rockville, Maryland

Jana M. Swenson, M.M.Sc.

Centers for Disease Control and Prevention
Atlanta, Georgia

Advisors

Richard H. Bell, Ph.D.

BAM, Ltd
Lathup Village, Michigan

Leslie V. Cunningham, Ph.D.

Baltimore, Maryland

George L. Evans, Ph.D.

White Haven, Pennsylvania

Mary Jane Ferraro, Ph.D., M.P.H.

Massachusetts General Hospital
Boston, Massachusetts

Frederic J. Marsik, Ph.D.

Food and Drug Administration
Rockville, Maryland

Mary E. Nichols

Remel
Lenexa, Kansas

Clyde Thornsberry, Ph.D.

Focus Technologies, Inc.
Franklin, Tennessee

Beth Ann Wise, M.T.(ASCP), M.S.Ed.
Staff Liaison

NCCLS
Wayne, Pennsylvania

Tracy A. Dooley, M.L.T.(ASCP)
Project Manager

NCCLS
Wayne, Pennsylvania

Patrice E. Polgar
Editor

NCCLS
Wayne, Pennsylvania

Donna M. Wilhelm
Assistant Editor

NCCLS
Wayne, Pennsylvania

Acknowledgements

The subcommittee acknowledges the efforts of the investigators who participated in studies upon which these guidelines are based in part. The support of the pharmaceutical industry and the four media manufacturers (Difco, BBL, Oxoid, and Hardy Diagnostics) that provided lots of Mueller-Hinton broth is gratefully acknowledged. The subcommittee thanks Trek Diagnostics for preparation of frozen trays for the investigations, and to Blaine Leppanen for keeping financial control during the investigation. The subcommittee also acknowledges the support of the Alberta Research Council, BD Diagnostic Systems, and Dade Behring Microscan for cation analyses, and Thea Heffner of the University of Alberta Hospital for statistical analysis of data.

Active Membership (as of 1 October 2001)

Sustaining Members

Abbott Laboratories
 American Association for
 Clinical Chemistry
 Bayer Corporation
 Beckman Coulter, Inc.
 BD and Company
 bioMérieux, Inc.
 CLMA
 College of American Pathologists
 GlaxoSmithKline
 Nippon Becton Dickinson Co., Ltd.
 Ortho-Clinical Diagnostics, Inc.
 Pfizer Inc
 Roche Diagnostics, Inc.

Professional Members

AISAR-Associazione Italiana per lo
 Studio degli
 American Academy of Family
 Physicians
 American Association for
 Clinical Chemistry
 American Association for
 Respiratory Care
 American Chemical Society
 American Medical Technologists
 American Public Health Association
 American Society for Clinical
 Laboratory Science
 American Society of Hematology
 American Society for Microbiology
 American Society of
 Parasitologists, Inc.
 American Type Culture
 Collection, Inc.
 Asociacion de Laboratorios de Alta
 Complejidad
 Asociación Española Primera de
 Socorros (Uruguay)
 Asociacion Mexicana de
 Bioquímica Clínica A.C.
 Assn. of Public Health Laboratories
 Assoc. Micro. Clinici Italiani-
 A.M.C.L.I.
 Australasian Association of
 Clinical Biochemists
 British Society for Antimicrobial
 Chemotherapy
 CADIME-Camara De Instituciones
 De Diagnostico Medico
 Bayer Corporation – Tarrytown, NY

Canadian Society for Medical
 Laboratory Science—Société
 Canadienne de Science de
 Laboratoire Médical
 Canadian Society of Clinical
 Chemists
 Clinical Laboratory Management
 Association
 COLA
 College of American Pathologists
 College of Medical Laboratory
 Technologists of Ontario
 College of Physicians and
 Surgeons of Saskatchewan
 Fundación Bioquímica Argentina
 International Association of Medical
 Laboratory Technologists
 International Council for
 Standardization in Haematology
 International Federation of
 Clinical Chemistry
 Italian Society of Clinical
 Biochemistry
 Japan Society of Clinical Chemistry
 Japanese Committee for Clinical
 Laboratory Standards
 Joint Commission on Accreditation
 of Healthcare Organizations
 National Academy of Clinical
 Biochemistry
 National Society for
 Histotechnology, Inc.
 Ontario Medical Association
 Quality Management Program-
 Laboratory Service
 RCPA Quality Assurance Programs
 PTY Limited
 Sociedade Brasileira de Analises
 Clinicas
 Sociedade Brasileira de
 Patologia Clinica
 Sociedad Espanola de Bioquímica
 Clinica y Patologia Molecular
 Turkish Society of Microbiology

Government Members

Association of Public Health
 Laboratories
 Armed Forces Institute of Pathology
 BC Centre for Disease Control
 Centers for Disease Control and
 Prevention
 Centers for Medicare & Medicaid
 Services/CLIA Program

Centers for Medicare & Medicaid
 Services
 Chinese Committee for Clinical
 Laboratory Standards
 Commonwealth of Pennsylvania
 Bureau of Laboratories
 Department of Veterans Affairs
 Deutsches Institut für Normung
 (DIN)
 FDA Center for Devices and
 Radiological Health
 FDA Center for Veterinary
 Medicine
 FDA Division of Anti-Infective
 Drug Products
 Iowa State Hygienic Laboratory
 Massachusetts Department of
 Public Health Laboratories
 National Association of Testing
 Authorities – Australia
 National Center of Infectious
 and Parasitic Diseases (Bulgaria)
 National Institute of Standards
 and Technology
 Ohio Department of Health
 Ontario Ministry of Health
 Saskatchewan Health-Provincial
 Laboratory
 Scientific Institute of Public Health;
 Belgium Ministry of Social
 Affairs, Public Health and the
 Environment
 South African Institute for Medical
 Research
 Swedish Institute for Infectious
 Disease Control
 Thailand Department of Medical
 Sciences

Industry Members

AB Biodisk
 Abbott Laboratories
 Abbott Laboratories, MediSense
 Products
 Accumetrics, Inc.
 Agilent Technologies, Inc.
 Ammirati Regulatory Consulting
 Anaerobe Systems
 Assessor
 AstraZeneca
 Aventis
 Avocet Medical, Inc.
 Axis-Shield PoC AS
 Bayer Corporation – Elkhart, IN

Bayer Corporation – West Haven, CT	Home Diagnostics, Inc.	Trek Diagnostic Systems, Inc.
Bayer Medical Ltd.	Immunicon Corporation	Vetoquinol S.A.
BD	I-STAT Corporation	Visible Genetics, Inc.
BD Biosciences – San Jose, CA	International Technidyne Corporation	Vysis, Inc.
BD Consumer Products	Kendall Sherwood-Davis & Geck	Wallac Oy
BD Diagnostic Systems	LAB-Interlink, Inc.	Wyeth-Ayerst
BD Italia S.P.A.	Labtest Diagnostica S.A.	Xyletech Systems, Inc.
BD VACUTAINER Systems	LifeScan, Inc. (a Johnson & Johnson Company)	YD Consultant
Beckman Coulter, Inc.	Lilly Research Laboratories	YD Diagnostics (Seoul, Korea)
Beckman Coulter, Inc. Primary Care Diagnostics	Medical Device Consultants, Inc.	
Beckman Coulter K.K. (Japan)	Medtronic, Inc.	Trade Associations
Bio-Development SRL	Merck & Company, Inc.	AdvaMed
Bio-Inova Life Sciences International	mvi Sciences (MA)	Association of Medical Diagnostic Manufacturers
Bio-Inova Life Sciences North America	Nabi	Japan Association Clinical Reagents Ind. (Tokyo, Japan)
BioMedia Laboratories Sdn Bhd	Neometrics is.	Medical Industry Association of Australia
bioMérieux, Inc.	Nichols Institute Diagnostics (Div. of Quest Diagnostics, Inc.)	
Biometry Consultants	Nissui Pharmaceutical Co., Ltd.	Associate Active Members
Bio-Rad Laboratories, Inc.	Nippon Becton Dickinson Co., Ltd.	20 th Medical Group (SC)
Bio-Rad Laboratories, Inc. - France	Norfolk Associates, Inc.	67 th CSH Wuerzburg, GE (NY)
Biotest AG	Organon Teknika Corporation	Academisch Ziekenhuis-VUB (Belgium)
Bristol-Myers Squibb Company	Ortho-Clinical Diagnostics, Inc. (Raritan, NJ)	Acadiana Medical Laboratories, LTD (LA)
Canadian External Quality Assessment Laboratory	Ortho-Clinical Diagnostics, Inc. (Rochester, NY)	Adena Regional Medical Center (OH)
Capital Management Consulting, Inc.	Oxoid Inc.	Advocate Laboratories (IL)
Checkpoint Development Inc.	Pfizer Inc	The Aga Khan Hospital & Medical College, Karachi (Pakistan)
Clinical Design Group Inc.	Pharmacia Corporation	Akershus Central Hospital and AFA (Norway)
Clinical Laboratory Improvement Consultants	Powers Consulting Services	Albany Medical Center Hospital (NY)
COBE Laboratories, Inc.	Premier Inc.	Albemarle Hospital (NC)
Community Medical Center (NJ)	Procter & Gamble Pharmaceuticals, Inc.	Allegheny General Hospital (PA)
Control Lab (Brazil)	The Product Development Group	Allegheny University of the Health Sciences (PA)
Copan Diagnostics Inc.	Quintiles, Inc.	Allina Laboratories (MN)
Cosmetic Ingredient Review	Radiometer America, Inc.	Alton Ochsner Medical Foundation (LA)
Cubist Pharmaceuticals	Radiometer Medical A/S	American Medical Laboratories (VA)
Cytometrics, Inc.	David G. Rhoads Associates, Inc.	Arkansas Department of Health
Dade Behring Inc. - Deerfield, IL	Roche Diagnostics GmbH	ARUP Laboratories (UT)
Dade Behring Inc. - Glasgow, DE	Roche Diagnostics, Inc.	ARUP at University Hospital (UT)
Dade Behring Inc. - Marburg, Germany	Roche Laboratories (Div. Hoffmann-La Roche Inc.)	Armed Forces Research Institute of Medical Science (APO, AP)
Dade Behring Inc. - Sacramento, CA	The R.W. Johnson	Aurora Consolidated Laboratories (WI)
Dade Behring Inc. - San Jose, CA	Pharmaceutical Research Institute	Azienda Ospedale Di Lecco (Italy)
DAKO A/S	Sarstedt, Inc.	Bay Medical Center (MI)
Diagnostic Products Corporation	Carl Schaper	Baystate Medical Center (MA)
Eiken Chemical Company, Ltd.	Schering Corporation	
Enterprise Analysis Corporation	Schleicher & Schuell, Inc.	
Fort Dodge Animal Health	Second Opinion	
General Hospital Vienna (Austria)	Showa Yakuhin Kako Company, Ltd.	
Gen-Probe	Streck Laboratories, Inc.	
GlaxoSmithKline	SurroMed, Inc.	
Greiner Bio-One Inc.	Sysmex Corporation (Japan)	
Health Systems Concepts, Inc.	Sysmex Corporation (Long Grove, IL)	
Helena Laboratories	The Toledo Hospital (OH)	

Bbaguas Duzen Laboratories (Turkey)
 Bo Ali Hospital (Iran)
 Bonnyville Health Center (Alberta, Canada)
 British Columbia Cancer Agency (Vancouver, BC, Canada)
 Broward General Medical Center (FL)
 Calgary Laboratory Services
 Carilion Consolidated Laboratory (VA)
 Carolinas Medical Center (NC)
 Cathay General Hospital (Taiwan)
 CB Healthcare Complex (Sydney, NS, Canada)
 Central Texas Veterans Health Care System
 Centre Hospitalier Regional del la Citadelle (Belgium)
 Centro Diagnostico Italiano (Milano, Italy)
 Champlain Valley Physicians Hospital (NY)
 Chang Gung Memorial Hospital (Taiwan)
 Children's Hospital (LA)
 Children's Hospital (NE)
 Children's Hospital & Clinics (MN)
 Children's Hospital King's Daughters (VA)
 Children's Hospital Medical Center (Akron, OH)
 Children's Hospital of Philadelphia (PA)
 Clarian Health–Methodist Hospital (IN)
 Clendo Lab (Puerto Rico)
 CLSI Laboratories (PA)
 Commonwealth of Kentucky
 CompuNet Clinical Laboratories (OH)
 Covance Central Laboratory Services (IN)
 Danville Regional Medical Center (VA)
 DeKalb Medical Center (GA)
 Delaware Public Health Laboratory
 Department of Health & Community Services (New Brunswick, Canada)
 DesPeres Hospital (MO)
 Detroit Health Department (MI)
 Diagnostic Laboratory Services, Inc. (HI)
 Duke University Medical Center (NC)
 Durham Regional Hospital (NC)
 Dynacare Laboratories - Eastern Region (Ottawa, ON, Canada)
 Dynacare Memorial Hermann Laboratory Services (TX)
 E.A. Conway Medical Center (LA)
 Eastern Maine Medical Center
 East Side Clinical Laboratory (RI)
 Elyria Memorial Hospital (OH)
 Emory University Hospital (GA)
 Esoterix Center for Infectious Disease (TX)
 Fairfax Hospital (VA)
 Fairview-University Medical Center (MN)
 Florida Hospital East Orlando
 Foothills Hospital (Calgary, AB, Canada)
 Fort St. John General Hospital (Fort St. John, BC, Canada)
 Fox Chase Cancer Center (PA)
 Franklin Square Hospital Center (MD)
 Fresenius Medical Care/Spectra East (NJ)
 Fresno Community Hospital and Medical Center
 Frye Regional Medical Center (NC)
 Gambro Healthcare Laboratory Services (FL)
 GDS Technology, Inc (IN)
 Geisinger Medical Center (PA)
 Grady Memorial Hospital (GA)
 Guthrie Clinic Laboratories (PA)
 Hahnemann University Hospital (PA)
 Harris Methodist Erath County (TX)
 Harris Methodist Fort Worth (TX)
 Hartford Hospital (CT)
 Headwaters Health Authority (Alberta, Canada)
 Health Network Lab (PA)
 Health Partners Laboratories (VA)
 Health Sciences Centre (Winnipeg, MB, Canada)
 Heartland Health System (MO)
 Highlands Regional Medical Center (FL)
 Hoag Memorial Hospital Presbyterian (CA)
 Holmes Regional Medical Center (FL)
 Holy Spirit Hospital (PA)
 Holzer Medical Center (OH)
 Hospital for Sick Children (Toronto, ON, Canada)
 Hospital Israelita Albert Einstein (Brazil)
 Hospital Sousa Martins (Portugal)
 Hotel Dieu Hospital (Windsor, ON, Canada)
 Huddinge University Hospital (Sweden)
 Hurley Medical Center (MI)
 Indiana State Board of Health
 Indiana University
 Instituto Scientifico HS. Raffaele (Italy)
 International Health Management Associates, Inc. (IL)
 Jackson Memorial Hospital (FL)
 Jersey Shore Medical Center (NJ)
 John F. Kennedy Medical Center (NJ)
 John Peter Smith Hospital (TX)
 Kadlec Medical Center (WA)
 Kaiser Permanente Medical Care (CA)
 Kaiser Permanente (MD)
 Kantonsspital (Switzerland)
 Kenora-Rainy River Regional Laboratory Program (Ontario, Canada)
 Kern Medical Center (CA)
 Kimball Medical Center (NJ)
 King Fahad National Guard Hospital (Saudi Arabia)
 King Faisal Specialist Hospital (Saudi Arabia)
 King Khalid National Guard Hospital (Saudi Arabia)
 King's Daughters Medical Center (KY)
 Klinični Center (Slovenia)
 LabCorp (NC)
 Laboratories at Bonfils (CO)
 Laboratoire de Santé Publique du Quebec (Canada)
 Laboratório Fleury S/C Ltda. (Brazil)
 Laboratory Corporation of America (NJ)
 Laboratory Corporation of America (MO)
 LAC and USC Healthcare Network (CA)
 Lakeland Regional Medical Center (FL)
 Lancaster General Hospital (PA)
 Langley Air Force Base (VA)
 LeBonheur Children's Medical Center (TN)
 Lewis-Gale Medical Center (VA)
 Libero Instituto Univ. Campus BioMedico (Italy)
 Long Beach Memorial Medical Center (CA)
 Louisiana State University Medical Center

Maccabi Medical Care and Health Fund (Israel)	North Shore – Long Island Jewish Health System Laboratories (NY)	St. Boniface General Hospital (Winnipeg, Canada)
Magee Womens Hospital (PA)	Northridge Hospital Medical Center (CA)	St. Elizabeth Hospital (NJ)
Manitoba Health (Winnipeg, Canada)	Northwestern Memorial Hospital (IL)	St-Eustache Hospital (Quebec, Canada)
Martin Luther King/Drew Medical Center (CA)	Ohio Valley Medical Center (WV)	St. John Hospital and Medical Center (MI)
Massachusetts General Hospital (Microbiology Laboratory)	O.L. Vrouwziekenhuis (Belgium)	St. John Regional Hospital (St. John, NB, Canada)
MDS Metro Laboratory Services (Burnaby, BC, Canada)	Ordre professionnel des technologistes médicaux du Québec	St. Joseph Hospital (NE)
Medical College of Virginia Hospital	Ospedali Riuniti (Italy)	St. Joseph’s Hospital – Marshfield Clinic (WI)
Medicare/Medicaid Certification, State of North Carolina	The Ottawa Hospital (Ottawa, ON, Canada)	St. Joseph’s Medical Center (CA)
Memorial Medical Center (IL)	Our Lady of Lourdes Hospital (NJ)	St. Luke’s Regional Medical Center (IA)
Memorial Medical Center (LA)	Our Lady of the Resurrection Medical Center (IL)	St. Mark’s Hospital (UT)
Jefferson Davis Hwy	Pathology and Cytology Laboratories, Inc. (KY)	St. Mary Medical Center (IN)
Memorial Medical Center (LA)	The Permanente Medical Group (CA)	St. Mary of the Plains Hospital (TX)
Napoleon Avenue	Piedmont Hospital (GA)	St. Mary’s Hospital & Medical Center (CO)
Mescalero Indian Hospital (NM)	Pocono Hospital (PA)	St. Paul’s Hospital (Vancouver, BC, Montreal)
Methodist Hospitals of Memphis (TN)	Polly Ryon Memorial Hospital (TX)	St. Vincent Medical Center (CA)
MetroHealth Medical Center (OH)	Presbyterian Hospital of Dallas (TX)	Ste. Justine Hospital (Montreal, PQ, Canada)
Michigan Department of Community Health	Prodia Clinical Laboratory (Indonesia)	Salina Regional Health Center (KS)
Monmouth Medical Center (NJ)	Providence Health System (OR)	San Francisco General Hospital (CA)
Monte Tabor – Centro Italo - Brasileiro de Promocao (Brazil)	Providence Seattle Medical Center (WA)	Santa Cabrini Hospital (Montreal, PQ Canada)
Montreal Children’s Hospital (Canada)	Queen Elizabeth Hospital (Prince Edward Island, Canada)	Santa Clara Valley Medical Center (CA)
Montreal General Hospital (Canada)	Queensland Health Pathology Services (Australia)	Seoul Nat’l University Hospital (Korea)
Morton Plant Mease Health Care (FL)	Quest Diagnostics Incorporated (CA)	Shanghai Center for the Clinical Laboratory (China)
MRL Pharmaceutical Services, Inc. (VA)	Quintiles Laboratories, Ltd. (GA)	South Bend Medical Foundation (IN)
MRL Reference Laboratory (CA)	Reading Hospital and Medical Center (PA)	Southern California Permanente Medical Group
National Institutes of Health (MD)	Regions Hospital	South Western Area Pathology Service (Australia)
National University Hospital (Singapore)	Reid Hospital & Health Care Services (IN)	Southwest Texas Methodist Hospital
Naval Surface Warfare Center (IN)	Research Medical Center (MO)	Speciality Laboratories, Inc. (CA)
Nebraska Health System	Rex Healthcare (NC)	Stanford Hospital and Clinics (CA)
New Britain General Hospital (CT)	Rhode Island Department of Health Laboratories	State of Washington Department of Health
New England Fertility Institute (CT)	Riyadh Armed Forces Hospital (Saudi Arabia)	Stormont-Vail Regional Medical Center (KS)
New England Medical Center Hospital (MA)	Royal Columbian Hospital (New Westminster, BC, Canada)	Sun Health-Boswell Hospital (AZ)
New York Hospital Medical Center of Queens	Sacred Heart Hospital (MD)	Sunrise Hospital and Medical Center (NV)
New York State Department of Health	Saint Mary’s Regional Medical Center (NV)	T.A. Sourasky Medical Center (Israel)
NorDx (ME)	St. Alexius Medical Center (ND)	Tampa General Hospital (FL)
North Carolina State Laboratory of Public Health	St. Anthony Hospital (CO)	Temple University Hospital (PA)
Northern Indiana Education Foundation	St. Barnabas Medical Center (NJ)	

Tenet Odessa Regional Hospital (TX)	University of Colorado Health Science Center	VA Outpatient Clinic (OH)
The Toledo Hospital (OH)	University of Chicago Hospitals (IL)	Vejele Hospital (Denmark)
Touro Infirmary (LA)	University of Florida	Viridae Clinical Sciences, Inc. (Vancouver, BC, Canada)
Trident Regional Medical Center (SC)	University of Illinois at Chicago	Washoe Medical Center Laboratory (NV)
Tripler Army Medical Center (HI)	University of the Ryukyus (Japan)	Watson Clinic (FL)
Truman Medical Center (MO)	University of Texas M.D. Anderson Cancer Center	Wilford Hall Medical Center (TX)
UCSF Medical Center (CA)	University of Virginia Medical Center	William Beaumont Hospital (MI)
UNC Hospitals (NC)	University of Washington	Williamsburg Community Hospital (VA)
University Hospital (Gent) (Belgium)	UPMC Bedford Memorial (PA)	Winn Army Community Hospital (GA)
University Hospitals of Cleveland (OH)	UZ-KUL Medical Center (Belgium)	Wishard Memorial Hospital (IN)
The University Hospitals (OK)	VA (Dayton) Medical Center (OH)	Yonsei University College of Medicine (Korea)
University of Alabama-Birmingham Hospital	VA (Denver) Medical Center (CO)	York Hospital (PA)
University of Alberta Hospitals (Canada)	VA (San Diego) Medical Center (CA)	Zale Lipshy University Hospital (TX)
	VA (Tuskegee) Medical Center (AL)	

OFFICERS

F. Alan Andersen, Ph.D.,
President
Cosmetic Ingredient Review

Donna M. Meyer, Ph.D.,
President Elect
CHRISTUS Health

Emil Voelkert, Ph.D.
Secretary
Roche Diagnostics GmbH

Gerald A. Hoeltge, M.D.
Treasurer
The Cleveland Clinic Foundation

William F. Koch, Ph.D.,
Immediate Past President
National Institute of Standards and Technology

John V. Bergen, Ph.D.,
Executive Director

Susan Blonshine, RRT, RPFT,
FAARC
TechEd

Wayne Brinster
BD

Kurt H. Davis, FCSMLS, CAE
Canadian Society for Medical
Laboratory Science

Robert L. Habig, Ph.D.
Molecular Diagnostics, Inc.

Thomas L. Hearn, Ph.D.
Centers for Disease Control and
Prevention

Carolyn D. Jones, J.D., M.P.H.
AdvaMed

BOARD OF DIRECTORS

Tadashi Kawai, M.D., Ph.D.
International Clinical Pathology
Center

J. Stephen Kroger, M.D., FACP
COLA

Gary L. Myers, Ph.D.
Centers for Disease Control and
Prevention

Barbara G. Painter, Ph.D.
Bayer Corporation

Ann M. Willey, Ph.D., J.D.
New York State Department of
Health

Judith A. Yost, M.A., M.T.(ASCP)
Centers for Medicare & Medicaid
Services

Contents

Abstract	i
Committee Membership.....	v
Active Membership.....	vii
Foreword	xv
1 Introduction	1
2 Scope	1
3 Definitions.....	1
4 Manufacturers' Protocol for Testing Production Lots of Dehydrated Mueller-Hinton Broth... 2	
4.1 Preparing Control Cultures	2
4.2 Performing the Tests	3
4.3 Interpreting the Results.....	5
4.4 Label Statement	5
5 Determining the Ion Content for Production Lots of Dehydrated Mueller-Hinton Broth..... 5	
5.1 Aminoglycosides.....	5
5.2 Carbapenems.....	5
5.3 Quinolones	5
5.4 Tetracycline.....	6
5.5 Ion Effects on Other Antimicrobial Agents (Not Tested in This Document)	6
5.6 Cation Concentrations.....	6
6 Testing New Antimicrobial Agents with Production Lots of Dehydrated Mueller-Hinton Broth.....	6
References	7
Appendix. Methods for Ion Analysis of Dehydrated Mueller-Hinton Broth	8
Related NCCLS Publications	9

Foreword

This proposed guideline is the result of the effort of the NCCLS Subcommittee on Culture Media to establish a standard protocol by which manufacturers of dehydrated Mueller-Hinton broth (dMHB) may determine its performance characteristics. This guideline is limited to preparation of production lots of dMHB for use according to NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#). The objectives of the project were:

- to evaluate possible variations in performance among currently manufactured lots of dMHB that are prepared for use in fresh or frozen reference MIC microdilution trays;
- to select performance criteria that will guide manufacturers of dMHB to produce lots of media that give accurate and reproducible results for broth microdilution antimicrobial susceptibility tests; and
- to characterize and quantitate medium components (especially cations and anions) that may influence the activity of an antimicrobial agent when tested *in vitro*.

The subcommittee attempted to define the characteristics of acceptable lots of dMHB. Nine laboratories participated in the investigations. Two lots of dMHB that had been released for sale were obtained from four different manufacturers. The identities of medium lots were blinded and coded, and the source of the media was known neither to the investigators nor to the manufacturer of the frozen broth microdilution trays. The control organisms were supplied from a single lot through the auspices of the American Type Culture Collection. Technologists in the participating laboratories were provided with specific instructions on how to thaw, inoculate, incubate, and read the broth microdilution trays so that each laboratory was performing the same methodology.

Cation and anion content was determined on samples from each lot of dMHB by inductively coupled plasma (ICP) analysis in laboratories at the Alberta Research Council. Subsequently these same samples were analyzed for major cations and anions by two different methods (ICP and flame atomic absorption spectroscopy) at BD Diagnostic Systems and Dade Behring Microscan to check for comparability of results among different testing laboratories. Results of ion concentrations in the eight lots of dMHB were found to vary by less than 5% among the three testing laboratories.

The following variables were analyzed for each of the antimicrobial/organism combinations tested:

- performance of each lot compared to quality control criteria in NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#);
- precision of the MIC observations; and
- interlaboratory variations.

The variability of cation and anion content in the lots of dMHB between and among manufacturers make it difficult to establish a "golden pound" of dMHB. The results of these laboratory investigations indicate that manufacturers should ensure that their production lots of dMHB conform to the quality control performance criteria identified in NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#). Contents of ions that are known or have been determined to affect *in vitro* activity of antimicrobial agents should be adjusted so that performance of the medium is appropriate. The content of critically important ions should be specifically identified on a certificate of analysis that is available to the user of the dMHB.

Invitation for Participation in the Consensus Process

An important aspect of the development of this and all NCCLS documents should be emphasized, and that is the consensus process. Within the context and operation of NCCLS, the term “consensus” means more than agreement. In the context of document development, “consensus” is a process by which NCCLS, its members, and interested parties (1) have the opportunity to review and to comment on any NCCLS publication; and (2) are assured that their comments will be given serious, competent consideration. Any NCCLS document will evolve as will technology affecting laboratory or healthcare procedures, methods, and protocols; and therefore, is expected to undergo cycles of evaluation and modification.

The NCCLS Area Committee on Microbiology has attempted to engage the broadest possible worldwide representation in committee deliberations. Consequently, it is reasonable to expect that issues remain unresolved at the time of publication at the proposed level. The review and comment process is the mechanism for resolving such issues.

The NCCLS voluntary consensus process is dependent upon the expertise of worldwide reviewers whose comments add value to the effort. At the end of a 90-day comment period, each subcommittee is obligated to review all comments and to respond in writing to all which are substantive. Where appropriate, modifications will be made to the document, and all comments along with the subcommittee's responses will be included as an appendix to the document when it is published at the next consensus level.

Key Words

Antimicrobial agent, broth microdilution, dehydrated Mueller-Hinton broth, minimal inhibitory concentration (MIC), susceptibility.

Evaluation of Lots of Dehydrated Mueller-Hinton Broth for Antimicrobial Susceptibility Testing; Proposed Guideline

1 Introduction

NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#) recommends the use of Mueller-Hinton broth for routine susceptibility testing for the following reasons:

- Historically there is good batch-to-batch reproducibility of QC results;
- The base medium, like Mueller-Hinton agar, is low in known inhibitors of sulfonamides, trimethoprim, and tetracycline;
- Mueller-Hinton broth supports the growth of most nonfastidious pathogens; and
- A large body of data is available using this medium to establish breakpoints and reproducible QC limits.

Earlier data from multicenter studies with Mueller-Hinton agar and other culture media showed that many variables affect agar diffusion susceptibility results. It appears that the culture medium itself is a critical component of that variability. The development of new antimicrobial agents and the more extensive use of Mueller-Hinton broth in laboratory studies from multicenter clinical trials indicate that more objective evaluation of possible variability of Mueller-Hinton broth is required.

This document provides protocols for evaluating the performance characteristics of production lots of dehydrated Mueller-Hinton broth (dMHB). The document only addresses lots of dMHB used to prepare fresh or frozen microdilution trays used in the broth dilution susceptibility method as described in the most recent version of NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#). It does not evaluate any variability that might occur as a result of use in dried or lyophilized media that are manufactured in microdilution MIC trays.

2 Scope

This proposed guideline is the result of the effort of the NCCLS Subcommittee on Culture Media to establish a standard protocol by which manufacturers of dehydrated Mueller-Hinton broth (dMHB) may determine its performance characteristics. This guideline is limited to preparation of production lots of dMHB for use according to NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#).

This document only addresses lots dMHB used to prepare fresh or frozen microdilution trays in the broth dilution susceptibility method as described in the most recent edition of NCCLS documents [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#). It does not evaluate any variability that might occur as a result of use in dried or lyophilized media that are manufactured in microdilution MIC trays.

3 Definitions

Dehydrated Mueller-Hinton broth, n – A dry-form complex bacteriological medium produced by several commercial manufacturers which is used to prepare broth dilution antimicrobial susceptibility media for testing routine nonfastidious aerobic and facultative anaerobic bacterial organisms.

Inoculum, *n* – A substance or portion of a specimen, implanted in a culture medium.

Minimal inhibitory concentration (MIC), *n* – The lower concentration of an antimicrobial agent that prevents visible growth of a microorganism in an agar or broth dilution susceptibility test.

4 Manufacturers' Protocol for Testing Production Lots of Dehydrated Mueller-Hinton Broth

4.1 Preparing Control Cultures

4.1.1 Stock Cultures

Stock cultures are prepared from lyophilized cultures obtained from the American Type Culture Collection (ATCC®). Cultures should be reconstituted according to procedures obtained from the ATCC® and maintained in such a way that selection of genetic variants is minimized. The cultures required for the purposes of this protocol are:

<i>Staphylococcus aureus</i>	ATCC®29213
<i>Staphylococcus aureus</i>	ATCC®43300
<i>Escherichia coli</i>	ATCC®25922
<i>Pseudomonas aeruginosa</i>	ATCC®27853
<i>Enterococcus faecalis</i>	ATCC®29212
<i>Streptococcus pneumoniae</i>	ATCC®49619

4.1.2 Preparing the Initial Stock Cultures

Using a sterile loop or swab, inoculate two to three plates of soybean-casein digest agar (tryptic soy agar; [TSA]) containing 5% sheep blood with each control culture listed in Section 4.1.1. Streak each plate to obtain isolated colonies. Incubate the plates for 18 to 24 hours at 35 ± 1 °C in ambient air, except for *S. pneumoniae* which should be incubated in an atmosphere of 5 to 7% CO₂.

4.1.3 Preparing Frozen Stock Cultures

After incubation, check for purity and harvest the entire growth from each set of plates and suspend it in soybean-casein digest broth (tryptic soy broth [TSB]) containing 15% glycerol. To prepare the TSB, add the manufacturer's recommended amount of TSB medium to make one liter to 500 mL of deionized water and add 150 mL of glycerol. Adjust the volume to 1 L, mix well, and sterilize it at 121 °C for 15 minutes. The bacterial suspension may be adjusted to the turbidity of a 0.5 McFarland standard (about 1 to 2 x 10⁸ CFU/mL), if a known viability is desired. Dispense at least 0.5 mL of the suspension into small, sterile vials. Store the vials at -60 °C, or a lower temperature. With this method, cultures should be viable for at least one year. Other methods of preparing stock cultures may be used if they provide adequate viability and stability. Periodically renew stock cultures from fresh lyophilized cultures obtained from ATCC®.

4.1.4 Preparing the Stock Culture Test Inoculum

The day before the inoculation of the plates (day 1), thaw a vial of each of the control cultures that will be needed. Inoculate each culture onto a plate of TSA with 5% sheep blood and incubate it for 18 to 24 hours (time is not critical for this step) at 35 °C in ambient air (5 to 7% CO₂ for *S. pneumoniae* ATCC® 49619). After incubation, check for purity. If satisfactory, growth from these plates is then used to prepare standardized inocula, as described in NCCLS document M7—*Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically*.

4.2 Performing the Tests

4.2.1 Preparing the Trays

Testing of each antimicrobial - organism combination should be performed to conform to the QC ranges indicated below. The minimum and maximum concentration on each tray should bracket the quality control limit range by at least two doubling dilutions beyond each limit. The following organisms should be tested against these agents using twofold dilutions and must fall within the current quality control limit ranges published in the most current edition of NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#) (and/or NCCLS document [M100—Performance Standards for Antimicrobial Susceptibility Testing](#)). The following values are included in the most current edition of NCCLS document [M100—Performance Standards for Antimicrobial Susceptibility Testing](#).

Quality Control Strain	Antimicrobial Agent	Quality Control Range (µg/mL)
<i>Pseudomonas aeruginosa</i> ATCC®27853	Tobramycin	0.25 – 1.0
	Imipenem	1.0 – 4.0
	Ciprofloxacin	0.25 – 1.0
<i>Escherichia coli</i> ATCC®25922	Ampicillin	2.0 – 8.0
	Cefotaxime	0.03 – 0.12
	Tetracycline	0.5. – 2.0
	Aztreonam	0.06 – 0.25
<i>Staphylococcus aureus</i> ATCC®29213	Clindamycin	0.06 – 0.25
	Azithromycin	0.5 – 2.0
	Oxacillin	0.12 – 0.5
	Erythromycin	0.25 – 1.0
<i>Enterococcus faecalis</i> ATCC®29212	Trimethoprim-sulfamethoxazole	≤ 0.5/9.5
	Vancomycin	1.0 – 4.0
<i>Staphylococcus aureus</i> ATCC®43300	Oxacillin	1.0 – 4.0
<i>Streptococcus pneumoniae</i> ATCC®49619	Penicillin	0.25 – 1.0
	Trimethoprim-sulfamethoxazole	0.12/2.4 – 1/19

4.2.2 Preparing Stock Antimicrobial Solutions

Stock solutions of antimicrobial agents should be prepared according to the methodology identified in NCCLS document [M100—Performance Standards for Antimicrobial Susceptibility Testing](#) (M7, Table 4). Final concentrations of antimicrobial agent should be prepared in the rehydrated Mueller-Hinton broth, adjusting cations if necessary. Dilutions may be prepared by making serial two/fold dilutions of the stock or by using the procedure described in [M100—Performance Standards for Antimicrobial Susceptibility Testing](#) (M7, Table 6).

- The final concentrations in each well of the trays should be contained in a total volume of 100 µL, and each tray should include a growth and negative control well.

- The addition of the final inoculum from a microdilution inoculator will not appreciably alter the final concentrations of antimicrobial agents if the volume of inoculum delivered is $\leq 10 \mu\text{L}$.

4.2.3 Preparing the Test Inoculum

A single inoculum for each quality control strain should be prepared using the direct colony suspension method as described in the most current version of NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#).

- Using sterile saline, adjust the turbidity of the suspension either visually or with a nephelometer or spectrophotometer to equal a 0.5 McFarland suspension.
- Within 15 minutes of preparation, dilute the adjusted inoculum with sterile saline so that, after inoculation, each tube or well contains approximately 5×10^5 CFU/mL. The dilution required to obtain this final inoculum varies according to the method of delivery of the inoculum to the individual wells or tubes and must be calculated for each situation. For microdilution tests the exact inoculum volume delivered to the wells must be known to make this calculation. For example, when the volume of medium in the well is 0.1 mL and the inoculum volume is 0.01 mL, then the 0.5 McFarland suspension ($\sim 1 \times 10^8$ CFU/mL) should be diluted 1:20 to yield 5×10^6 CFU/mL. When 0.01 mL of this dilution is inoculated into the broth, the final test concentration of bacteria will be approximately 5×10^5 CFU/mL (or 5×10^4 CFU/well).

4.2.4 Performing Colony Counts

Perform colony counts on the inoculated plate to verify that the final inoculum is approximately 5×10^5 CFU/mL as per NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#). ([Section 7.3.1\[4\]](#)).

- Immediately after inoculating, transfer 0.01 mL from the growth control well into a tube containing 10 mL of sterile 0.9% saline.
- Transfer a 0.1-mL sample onto the surface of a suitable agar medium, and spread the liquid over the entire plate surface.
- After incubation the presence of 50 colonies indicates an inoculum density of 5×10^5 CFU/mL.

4.2.5 Incubating the Microdilution Trays

Trays should be incubated in ambient air at 35 °C:

- Incubate *S. pneumoniae* ATCC[®] 49619 and *S. aureus* ATCC[®] 43300 (for oxacillin) for 24 hours.
- Incubate all other strains for 16 to 18 hours.

4.2.6 Reading the Microdilution Trays

After incubation, trays should be removed from the incubator and read within one hour. The endpoint (MIC) is defined as the lowest concentration that completely inhibits the growth of the organism. For trimethoprim-sulfamethoxazole (TMP-SMX), the endpoint (MIC) is defined as 80% or greater reduction of growth compared to the control well (no antimicrobial agent).

4.3 Interpreting the Results

The MIC for each antimicrobial-organism concentration should be within the quality control limits identified in the [M7 section of NCCLS document M100—Performance Standards for Antimicrobial Susceptibility Testing](#) for that combination. If the susceptibility test results are within MIC quality control limits for each antimicrobial agent-organism combination, the dMHB lot is acceptable. If any organism-antimicrobial agent susceptibility result falls outside the established quality control MIC limits, corrective action must be taken. That antimicrobial agent-organism test should be repeated, so that the test results fall within the quality control limits at least 95% of the time (i.e., 19 of 20 times).

4.4 Label Statement

If all performance criteria for all organism-antimicrobial combinations are within acceptable limits, the manufacturer may indicate for that production lot that the dMHB (lot number) conforms to the performance criteria laid out in NCCLS document M32 for antimicrobial susceptibility testing. For standard production lots of dMHB, the broth prepared from the dehydrated product should contain no greater than 25 mg/L of calcium, 12.5 mg/L of magnesium, and must have less than 3 mg/L of zinc ions. A certificate of analysis should be available to the user of dMHB that specifies the concentrations of these ions in the lot of dehydrated medium and indicates if additional calcium and/or magnesium is required to achieve the concentrations specified in NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#)

The label or package insert statement should state:

"This lot of dehydrated Mueller-Hinton broth has been tested according to, and meets quality control criteria as stated in NCCLS document M32."

5 Determining the Ion Content for Production Lots of Dehydrated Mueller-Hinton Broth

Ion content for major cations and anions known or suspected as having an effect on antimicrobial-organism interactions should be tested by inductively coupled plasma analysis (ICP) or flame atomic absorption spectroscopy (FAAS) using a recognized standard method that has been validated. A brief summary description of these methods with sensitivity and detection limit data is given in the appendix. Available data indicates that interactions between ions and the following antimicrobial agents are known or may cause changes in broth dilution antimicrobial susceptibility test results.

5.1 Aminoglycosides

Suggested levels of calcium (20 to 25 µg/mL) and magnesium (10 to 12.5 µg/mL) are based on studies ^{1,2} comparing Mueller-Hinton agar and broth dilution for clinical strains of *P. aeruginosa*.

5.2 Carbapenems

The concentration of zinc should be below 3 µg/mL to avoid false resistance interpretations.³ This is known to be true for imipenem and may apply to other carbapenems.

5.3 Quinolones

Data from studies of quinolone activity in human urine suggest that decreased quinolone activity may occur when magnesium concentrations are in the range of 8 to 10 mM (> 100 to 150 µg/mL) in Mueller-

Hinton medium.⁴ Other data indicate that 35 to 60 µg/mL of magnesium cause a reduction in zone diameters and an increase in MICs for several species of bacteria.^{5,6}

5.4 Tetracycline

Mueller-Hinton broth supplemented to contain 50 µg/mL of calcium and 25 µg/mL of magnesium has been shown to increase MICs of *E. coli*, *P. aeruginosa*, and other species of *Pseudomonas* by 2- to 32-fold.⁷

5.5 Ion Effects on Other Antimicrobial Agents (Not Tested in This Document)

Although ion effects on other antimicrobial agents not included in this document are known to affect susceptibility results, they should be considered for Mueller-Hinton broth dilution susceptibility tests. (These include daptomycin,⁸⁻¹⁰ polymyxin,⁷ and mecillinam.¹¹) Manufacturers and others should refer to these references and to NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#) for appropriate instructions on preparation and testing.

5.6 Cation Concentrations

If test results of production lots of dMHB do not conform to the quality control ranges identified in the [M7](#) section of [M100—Performance Standards for Antimicrobial Susceptibility Testing](#) and the ion concentrations of calcium, magnesium, and zinc are within those identified in [Section 4.4](#) of this document and in NCCLS document [M7—Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria that Grow Aerobically](#), concentrations of other ions or medium components should be checked to determine if they affect the results.

6 Testing New Antimicrobial Agents with Production Lots of Dehydrated Mueller-Hinton Broth

When *in vitro* data on new antimicrobial agents are being developed, dMHB production lots meeting the criteria in this document should be used to develop *in vitro* quality control parameters for these new agents. Testing of production lots of dMHB with new antimicrobial agents should follow the procedures outlined in this document. Ion content should be examined to determine if the new agent may be affected by specific cations or anions or concentrations of ions in the medium that differ from the ranges suggested in this guidance document. Other medium components that may affect *in vitro* quality control susceptibility tests that come to light during these investigations should be identified. Adjustments should be made as necessary to achieve stable, reproducible test results, and this information should be widely circulated to others involved in antimicrobial susceptibility testing and quality control including the NCCLS Subcommittees on Culture Media and Antimicrobial Susceptibility Testing.

References

- ¹ Barry AL, Miller GH, Thornsberry C, et al. Influence of cation supplements on activity of netilmicin against *Pseudomonas aeruginosa* *in vitro* and *in vivo*. *Antimicrob Agents Chemother.* 1987;31:1514-1518.
- ² Barry AL, Reller LB, Miller GH, et al. Revision of standards for adjusting the cation content of Mueller-Hinton broth for testing susceptibility of *Pseudomonas aeruginosa* to aminoglycosides. *J Clin Microbiol.* 1992;30:585-589.
- ³ Daly JS, Dodge RA, Glew RH, Soja DT, DeLuca BA, Hebert S. Effect of zinc concentration in Mueller-Hinton agar on susceptibility of *Pseudomonas aeruginosa* to imipenem. *J Clin Microbiol.* 1997;35:1027-1029.
- ⁴ Eliopoulos GM, Eliopoulos CT. Quinolone antimicrobial agents: Activity *in vitro*. In Wolfson JS, Hooper DC, eds. *Quinolone Antimicrobial Agents*. Washington, D.C.: American Society for Microbiology; 1989:3:35-70.
- ⁵ Auckenthaler R, Michea-Hamzeshpour M, Pechere JC. *In vitro* activity of newer quinolones against aerobic bacteria. *J Antimicrob Chemother.* 1986;17(Suppl. B):29-39.
- ⁶ Blaser J, Dudley MN, Gilbert D, Zinner SH. Influence of medium and method on the *in vitro* susceptibility of *Pseudomonas aeruginosa* and other bacteria to ciprofloxacin and enoxacin. *Antimicrob Agents Chemother.* 1986;29:927-929.
- ⁷ Fass RJ, Barnshaw J. Effect of divalent cation concentrations on the antibiotic susceptibilities of non-fermenters other than *Pseudomonas aeruginosa*. *Antimicrob Agents Chemother.* 1979;16:434-438.
- ⁸ Brown SD, Barry AL, Fuchs PC. *In vitro* activity of daptomycin (Cidecin) against contemporary Gram-positive clinical bacterial isolates from 11 North American medical centers (NAMC). *ECCMID.* 2000; Abstract P90:5/5.
- ⁹ Fuchs PC, Barry AL, Brown SD. Daptomycin susceptibility tests: interpretive criteria, quality control, and effect of calcium on *in vitro* tests. *Diag Microbiol Infect Dis.* 2000;38:51-58.
- ¹⁰ Barry AL, Fuchs PC, Brown SD. *In vitro* activities of daptomycin against 2,789 clinical isolates from 11 North American medical centers. *Antimicrob Agents Chemother.* 2001;45:1919-1922.
- ¹¹ Neu HC. Mecillinam, a novel penicillanic acid derivative with unusual activity against Gram-negative bacteria. *Antimicrob Agents Chemother.* 1976;9:793-799.

Appendix. Methods for Ion Analysis of Dehydrated Mueller-Hinton Broth

Flame Atomic Absorption Spectroscopy (FAAS). The instrument consists of an atomizer, burner, light source, wavelength selector, and detector. The sample is first atomized, and then burned. The light source is a hollow cathode lamp (HCL). The light from the HCL is passed through the flame, through a wavelength selector, and to the detector where the amount of light is measured. When a sample is introduced into the flame, analyte atoms will absorb the quantized light from the HCL and reach an excited state. When the atoms de-excite most of the re-emitted quantized light is radiated away from the monochromator and the detector, so the amount of light passing from the HCL through the flame is reduced. The reduction in light intensity, or absorbance, is proportional to the number of atoms of the element of interest in the flame and subsequently to the concentration of the sample introduced into the flame.

Inductively Coupled Plasma Spectroscopy (ICP). The instrument consists of an atomizer, plasma source, wavelength selector, and detector. The sample is mixed with argon gas, which is then ionized and forced to move in a circular path by a radio frequency generator. The resistance of the ions to the applied radio frequency results in extremely hot temperatures, up to 10,000 °K. The energy that results from heating and subsequent cooling is in the form of electromagnetic radiation, or light. The light from the plasma is passed through a wavelength selector and to the detector where the amount of light is measured. The increase in light intensity, or absorbance, is proportional to the number of atoms of the element of interest in the plasma and subsequently to the concentration of the sample introduced into the plasma.

Sensitivity, Detection Limit, and Accuracy. The sensitivity and detection limit of FAAS and ICP are affected by variables such as flame/plasma temperature, spectral bandwidth, and detector sensitivity. In the absence of sample matrix effects, sensitivity and detection limit are generally close to each other. For flame atomic absorption spectroscopy, sensitivities from about 3×10^{-4} to 20 $\mu\text{g/mL}$ are observed. For inductively coupled plasma spectroscopy, detection limits from about 3×10^{-6} to 1×10^{-3} $\mu\text{g/mL}$ are observed. The table lists some detection limits and sensitivities for FAAS and ICP. The accuracy of FAAS and ICP are generally less than 5%.

Table. Sensitivity and Detection Limits for FAAS and ICP. (Reprinted with permission from Morrison GH. *Critical Reviews in Analytical Chemistry*. 1969;(14):28A. ©CRC Press, Boca Raton, Florida.)

Element	FAAS Sensitivity ($\mu\text{g/mL}$)	ICP Detection Limit ($\mu\text{g/mL}$)
Calcium (Ca)	0.02	0.000001
Magnesium (Mg)	0.003	0.000003
Iron (Fe)	0.06	0.0005
Copper (Cu)	0.04	0.0004
Sodium (Na)	0.003	0.0002
Zinc (Zn)	0.009	0.0001

Related NCCLS Publications*

- M2-A7** **Performance Standards for Antimicrobial Disk Susceptibility Tests; Approved Standard—Seventh Edition (2000).** American National Standard.* This standard contains updated recommended techniques, interpretive criteria, and quality control parameters for disk susceptibility testing.
- M6-A** **Protocols for Evaluating Dehydrated Mueller-Hinton Agar; Approved Standard (1996).** This document provides procedures for evaluating production lots of Mueller-Hinton agar, and for the development and application of reference media.
- M7-A5** **Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria That Grow Aerobically; Approved Standard—Fifth Edition (2000).** American National Standard.* This standard provides updated reference methods for the determination of minimal inhibitory concentrations (MICs) for aerobic bacteria by broth macrodilution, broth microdilution, and agar dilution.
- M22-A2** **Quality Assurance for Commercially Prepared Microbiological Culture Media—Second Edition; Approved Standard (1996).** This document provides quality assurance procedures for manufacturers and users of prepared, ready-to-use microbiological culture media.
- M29-A2** **Protection of Laboratory Workers from Occupationally Acquired Infections—Second Edition; Approved Guideline (2001).** This document provides guidance on the risk of transmission of hepatitis viruses and human immunodeficiency viruses in any laboratory setting; specific precautions for preventing the laboratory transmission of blood-borne infection from laboratory instruments and materials; and recommendations for the management of blood-borne exposure.
- M31-A** **Performance Standards for Antimicrobial Disk and Dilution Susceptibility Tests for Bacteria Isolated from Animals; Approved Standard (1999).** Reference methods for the determination of minimum inhibitory concentration (MIC) and agar disk diffusion susceptibility tests with bacteria isolated from animals, broth macrodilution, broth microdilution, agar dilution, and agar disk diffusion methods.
- M100-S12** **Performance Standards for Antimicrobial Susceptibility Testing; Twelfth Informational Supplement (2002).** This document provides updated tables for the NCCLS antimicrobial susceptibility testing standards M2-A7 and M7-A5.

* Proposed- and tentative-level documents are being advanced through the NCCLS consensus process; therefore, readers should refer to the most recent editions.

NCCLS ▼ 940 West Valley Road ▼ Suite 1400 ▼ Wayne, PA 19087 ▼ USA ▼ PHONE 610.688.0100
FAX 610.688.0700 ▼ E-MAIL: exoffice@nccls.org ▼ WEBSITE: www.nccls.org ▼ ISBN 1-56238-447-3

